

219533

CERTIFICATE

of conformity
No. 911055



In compliance with the requirements for documentation specified in Annex III (art. 2) of the European Parliament and of the Council on the market of electrical equipment designed for use within certain voltage limits, this certificate applies to the electrical equipment:

ETO STERILIZER

Technical specification: **LOW VOLTAGE DIRECTIVE**
Technical file: **NSEPL/TCF/01**
Assessment Performed: **LOW VOLTAGE DIRECTIVE 2014/35/EU**
Category (Annex IV): **LVD (EN 60204-1)**
Applied Standards: **ISO-12100:2010, IEC 61326-1:2005**

produced by or for

NAT STEEL EQUIPMENT PVT. LTD.
G.D. Ambekar Marg, Naigaum Road, Dadar (E), Mumbai - 400014,
Maharashtra, India.

and produced in the manufacturing plant(s)
NAT STEEL EQUIPMENT PVT. LTD.
G.D. Ambekar Marg, Naigaum Road, Dadar (E), Mumbai - 400014,
Maharashtra, India.

SEE ENCLOSURE

This certificate attests that all provisions described in relevant parts of the standard

EN 60204-1

are addressed in the accompanying product documentation. The CE mark as shown below can be affixed, under the responsibility of the manufacturer, after the completion of EC declaration of conformity and compliance all the relevant directives.

This certificate was issued on **30 August 2022** and is valid until **29 August 2027** and is based on the evaluation of the technical file of the equipment/machinery device. The voluntary certification does not imply an assessment of the production and it does not permit the use of a mark of conformity or of a safety mark of the LL-C (Certification). The holder of this certificate may use this certificate together with his EC declaration of conformity. This is not NB statement.

R. N. Kuwad

Prague, 30 August 2022

Rohit Kuwad
Deputy of Head of CB



validity code: **337B5AFD-53E**

Check the validity of this certificate using this code at www.ll-c.info

Annex to certificate

No. 911055

NAT STEEL EQUIPMENT PVT. LTD.

G.D. Ambekar Marg, Naigaum Road, Dadar (E), Mumbai - 400014, Maharashtra, India.

VAT: 15520



Standard

2014/35/EU - LVD - EN 60204-1

Parameters

No.	Name product	Technical specification	Model	Standard
1	ETO STERILIZER		1-2SR/SH/DV/FA ; 2-4SR/SH/DV/FA	ISO-13849-1:2015, ISO-12100:2010, IEC 61010-2-040:2005 & IEC 61010-1:2010-06, IEC 61326-1:2005.
2	ETO STERILIZER		3-6SR/SH/DV/FA ; 4-6SR/SH/DV/FA/SPL	ISO-13849- 1:2015, ISO-12100:2010, IEC 61010-2-040:2005 & IEC 61010-1:2010-06, IEC 61326-1:2005.
3	ETO STERILIZER		5-9SR/HW/WV/FA ; 6-16SR/HW/WV/FA/SPL	ISO-13849- 1:2015, ISO-12100:2010, IEC 61010-2-040:2005 & IEC 61010-1: 2010-06, IEC 61326-1:2005.
4	ETO STERILIZER		7-16SR/HW/WV/FA ; 8-32SR/HW/WV/FA	ISO-13849- 1:2015, ISO-12100:2010, IEC 61010-2-040:2005 & IEC 61010-1:2010-06, IEC 61326-1:2005.
5	ETO STERILIZER		9-40SR/HW/WV/FA	ISO-13849- 1:2015, ISO-12100:2010, IEC 61010-2-040:2005 & IEC 61010-1: 2010-06, IEC 61326-1:2005.

R. N. Kuwad

Approved by

THIS ANNEX IS NOT VALID WITHOUT MAIN CERTIFICATE

Verification code: **337B5AFD-53E**

Check the validity using this code at www.ll-c.info

LL-C (Certification) Czech Republic a.s. | Pobrezni 620/3, 186 00 Praha 8